

Quality Control Testing Official Report

BD-2504

Sample ID: G5B0307-06Matrix: Extracts & ConcentratesTest ID: 5035452Source ID:

Date Sampled: 02/26/25

Date Accepted: 02/26/25

Jefferson E-Commerce

rebekah@roguecloudco.com

Results at a Glance

Total THC : 68.46 %

Total CBD : <LOQ (0.0431%) %



Nolan Mundie Lab Director - 3/3/2025

Page 1 of 4



Quality Control Testing Official Report

BD-2504

Sample ID: G5B0307-06 Matr Test ID: 5035452 Source ID:

Date Sampled: 02/26/25

Matrix: Extracts & Concentrates

Date Accepted: 02/26/25

Jefferson E-Commerce

rebekah@roguecloudco.com

Date/Time Extracted: 02/28/25 10:15				Analysis Method/SOP: 215 Batch Identification: 250905		
annabinoids	LOQ (%)	% by Wt.	mg/g	Cannabinoids Profile		
Total THC	0.0005	68.46	684.6			
Total CBD	0.0431	< LOQ	< LOQ			
THCA	0.0005	78.06	780.6	-2.9		
delta 9-THC	0.0005	< LOQ	< LOQ			
delta 8-THC	0.0934	< LOQ	< LOQ			
THCV	0.1052	< LOQ	< LOQ			
THCVA	0.0392	< LOQ	< LOQ			
CBD	0.0005	< LOQ	< LOQ			
CBDA	0.0005	< LOQ	< LOQ			
CBDV	0.1040	< LOQ	< LOQ	THCA 78.1 CBGA 2.9		
CBDVA	0.0341	< LOQ	< LOQ	Total: 81.0		
CBN	0.0622	< LOQ	< LOQ			
CBG	0.0164	< LOQ	< LOQ			
CBGA	0.0164	2.930	29.3	78.1		
CBC	0.0186	< LOQ	< LOQ			
Total Cannabinoids 80.99		809.9				

Total THC = delta 9-THC + (THCA * 0.877) Total CBD = CBD + (CBDA * 0.877) Total CBG = CBG + (CBGA * 0.878) LOQ=Limit of Quantification, the lowest measurable concentration of an analyte. THCA, delta 9-THC, delta 8-THC, CBDA and CBD are accredited by TNI 2016 and ISO 17025



Nolan Mundie Lab Director - 3/3/2025



Quality Control Potency

Batch: 2509057 - 215-Concentrates

Blank(2509057-BLK2)									
Analyte	Result	LOQ	Units	%Recovery Limits	Extracted	Analyzed	Notes		
THCA	< LOQ	0.0003	%		02/28/25 10:15	03/03/25 13:17			
delta 9-THC	< LOQ	0.0003	%		02/28/25 10:15	03/03/25 13:17			
delta 8-THC	< LOQ	0.0566	%		02/28/25 10:15	03/03/25 13:17			
THCV	< LOQ	0.0638	%		02/28/25 10:15	03/03/25 13:17			
THCVA	< LOQ	0.0238	%		02/28/25 10:15	03/03/25 13:17			
CBD	< LOQ	0.0003	%		02/28/25 10:15	03/03/25 13:17			
CBDA	< LOQ	0.0003	%		02/28/25 10:15	03/03/25 13:17			
CBDV	< LOQ	0.0630	%		02/28/25 10:15	03/03/25 13:17			
CBDVA	< LOQ	0.0207	%		02/28/25 10:15	03/03/25 13:17			
CBN	< LOQ	0.0377	%		02/28/25 10:15	03/03/25 13:17			
CBG	< LOQ	0.0099	%		02/28/25 10:15	03/03/25 13:17			
CBGA	< LOQ	0.0099	%		02/28/25 10:15	03/03/25 13:17			
CBC	< LOQ	0.0113	%		02/28/25 10:15	03/03/25 13:17			
Reference(250	9057-SRM2)								
Analyte	% Recovery	LOQ	Units	%Recovery Limits	Extracted	Analyzed	Notes		

Analyte	% Recovery	LOQ	Units	%Recovery Limits	Extracted	Analyzed	Notes
THCA	100	0.00023	%	90-110	02/28/25 10:15	03/03/25 13:39	
delta 9-THC	102	0.00023	%	90-110	02/28/25 10:15	03/03/25 13:39	
delta 8-THC	96.6	0.0432	%	90-110	02/28/25 10:15	03/03/25 13:39	
CBD	96.9	0.00023	%	90-110	02/28/25 10:15	03/03/25 13:39	
CBDA	98.2	0.00023	%	90-110	02/28/25 10:15	03/03/25 13:39	



Muntin

Nolan Mundie Lab Director - 3/3/2025

Page 3 of 4



Notes and Definitions

Regulatory Compliance samples were collected onsite at facility according to SOP-402 and SOP-403 and following Sampling Plan FN117. Quality Control samples were tested as received. Results do not include uncertainty of measurements. Available upon request.

- ATM Non-cannabis matrix related interference or suppression of Internal standard
- BLI Baseline Interference Cannabinoid peak interference in chromatographic baseline affecting QC recovery .
- BLK Analyte detected in method blank, but not associated samples.
- BSH Blank Spike High Blank Spike recovery above method limit. no detections in samples.
- BSL Blank Spike Low Blank Spike recovery below lower method limit, analyte chromatography reviewed
- C manually for all samples.
- CBD Interference due to co-elution
- CV1 CBD matrix interference on GC Pest chromatography
- CV2 CCV was above acceptance criteria, Non-detect samples are considered acceptable.
- INF CCV was below acceptance criteria, sample still exceeds regulatory limit.
- ISH One or more QC falls outside acceptance criteria. Data entered into LIMS for informational purposes only.
- ISL Internal Standard concentration is above acceptance criteria.
- MSH Internal Standard concentration is below acceptance criteria.
- MSI Matrix Spike High Matrix Spike recovery above method limits.
- MSL Matrix Spike Interference Matrix spike source sample contains analyte hit above calibration affecting
- TPP recovery accuracy in Matrix Spike.
- U Matrix Spike Low Matrix Spike recovery below lower method limit, analyte chromatography reviewed manually for all samples.

Internal Standard concentration outside control limit due to matrix interference



Nolan Mundie Lab Director - 3/3/2025